

	Type	Hits	Search Text
1	BRS	213	(vector\$3 ADJ network ADJ analy\$4) SAME calibrat\$5
2	BRS	23	((vector\$2 ADJ network ADJ (analyzer OR analyser)) AND calibrat\$4 AND short AND open).clm.
3	BRS	8	((vector\$2 ADJ network ADJ (analyzer OR analyser)) AND calibrat\$4 AND short AND open) AND determin\$3 AND (unknown NEAR3 reflect\$5) AND match AND error AND coefficient AND standard

Calibration process for multiport network analyzer based on a 7-term process

Publication number: DE19918960

Publication date: 1999-11-11

Inventor: HEUERMAN HOLGER (DE); FABRY HANS-JOACHIM (DE); BALLMANN RALF (DE)

Applicant: HEUERMAN HOLGER (DE); FABRY HANS JOACHIM (DE); BALLMANN RALF (DE)

Classification:

- international: **G01R35/00; G01R27/28; G01R35/00; G01R27/00;**
(IPC1-7): G01R35/00; G01R27/28

- european: G01R35/00

Application number: DE19991018960 19990427

Priority number(s): DE19991018960 19990427; DE19981018877 19980428

Report a data error here

Abstract of **DE19918960**

Calibration process for an n port network analyzer by measurement of reflection and transmission parameters at n+1 different but sequential calibration standards between the measurement ports. Calibration standards are taken for known n-port, 2 port or up to n times single port arrangements. Calibration standards according to any of a number of known 7 port processes can be used.

Data supplied from the **esp@cenet** database - Worldwide

Calibrating vectorial network analyzer

Publication number: DE19918697

Publication date: 1999-11-18

Inventor: HEUERMAN HOLGER (DE); FABRY HANS-JOACHIM (DE); BALLMANN RALF (DE)

Applicant: HEUERMAN HOLGER (DE); FABRY HANS JOACHIM (DE); BALLMANN RALF (DE)

Classification:

- international: **G01R35/00; G01R35/00**; (IPC1-7): G01R35/00; G01R27/28

- european: G01R35/00C

Application number: DE19991018697 19990426

Priority number(s): DE19991018697 19990426; DE19981018878 19980428

Report a data error here

Abstract of **DE19918697**

All calibrating standards must consist of completely known n gates, n times input gates (n-gates consisting of input gates). At least one signal path of finite transmission damping as calibrating standard, must be connected between each measuring gate combination.

Data supplied from the **esp@cenet** database - Worldwide